



CERTIFICATE OF CONFORMITY

PRODUCT: SHER INDICATOR

ORDER No: 003512

LOT No: SHER16J1

DATE OF TEST: 05/09/2016

PRODUCTION DATE: 02/09/2016

EXPIRY DATE: 28/03/2018

SPECIFICATION: Passes Sensitivity Test
Using 4% Boric Acid, (pH 4) Orange / (pH 7.5) Green

This certificate is electronically generated and does not require a signature.

Date: 05/09/2016

This certificate must not be reproduced except in full.

BÜCHI Labortechnik AG
Meierseggstrasse 40
9230 Flawil
Switzerland
T +41 71 394 63 63
F +41 71 394 64 64
www.buchi.com