



## **CERTIFICATE OF CONFORMITY**

**PRODUCT:** SHER MIXED INDICATOR

**ORDER No:** 003512

**LOT No:** SHER22B1

**DATE OF TEST:** 07/02/2022

**PRODUCTION DATE:** 07/02/2022

**EXPIRY DATE:** 28/08/2023

**SPECIFICATION:** Passes Sensitivity Test  
Using 4% Boric Acid; (pH 4) Orange, (pH 7.5) Green

This certificate is electronically generated and does not require a signature.

**Date:** 07/02/2022

This certificate must not be reproduced except in full.

**BÜCHI Labortechnik AG**  
Meierseggstrasse 40  
9230 Flawil  
Switzerland  
T +41 71 394 63 63  
F +41 71 394 64 64  
[www.buchi.com](http://www.buchi.com)