



CERTIFICATE OF CONFORMITY

PRODUCT: SHER MIXED INDICATOR

ORDER No: 003512

LOT No: SHER19M1

DATE OF TEST: 12/12/2019

PRODUCTION DATE: 06/12/2019

EXPIRY DATE: 28/06/2021

SPECIFICATION: Passes Sensitivity Test
Using 4% Boric Acid; (pH 4) Orange, (pH 7.5) Green

This certificate is electronically generated and does not require a signature.

Date: 17/12/2019

This certificate must not be reproduced except in full.

BÜCHI Labortechnik AG
Meierseggrasse 40
9230 Flawil
Switzerland
T +41 71 394 63 63
F +41 71 394 64 64
www.buchi.com