

CERTIFICATE OF CONFORMITY

PRODUCT: SHER INDICATOR

ORDER No: 003512

LOT No: SHER16E1

DATE OF TEST: 01/06/2016

PRODUCTION DATE: 25/05/2016

EXPIRY DATE: 28/11/2017

SPECIFICATION: Passes Sensitivity Test

Using 4% Boric Acid, (pH 4) Orange / (pH 7.5) Green

This certificate is electronically generated and does not require a signature.

Date: 01/06/2016

This certificate must not be reproduced except in full.

BÜCHI Labortechnik AG Meierseggstrasse 40 9230 Flawil Switzerland T +41 71 394 63 63 F +41 71 394 64 64 www.buchi.com