



CERTIFICATE OF CONFORMITY

PRODUCT: SHER INDICATOR

ORDER No: 003512

LOT No: SHER17E1

DATE OF TEST: 11/05/2017

PRODUCTION DATE: 08/05/2017

EXPIRY DATE: 28/11/2018

SPECIFICATION: Passes Sensitivity Test
Using 4% Boric Acid, (pH 4) Orange / (pH 7.5) Green

This certificate is electronically generated and does not require a signature.

Date: 11/05/2017

This certificate must not be reproduced except in full.

BÜCHI Labortechnik AG
Meierseggstrasse 40
9230 Flawil
Switzerland
T +41 71 394 63 63
F +41 71 394 64 64
www.buchi.com