



## **CERTIFICATE OF CONFORMITY**

**PRODUCT:** SHER MIXED INDICATOR

**ORDER No:** 003512

**LOT No:** SHER18K1

**DATE OF TEST:** 22/10/2018

**PRODUCTION DATE:** 22/10/2018

**EXPIRY DATE:** 28/04/2020

**SPECIFICATION:** Passes Sensitivity Test  
Using 4% Boric Acid, (pH 4) Orange / (pH 7.5) Green

This certificate is electronically generated and does not require a signature.

**Date:** 22/10/2018

This certificate must not be reproduced except in full.

**BÜCHI Labortechnik AG**  
Meierseggstrasse 40  
9230 Flawil  
Switzerland  
T +41 71 394 63 63  
F +41 71 394 64 64  
[www.buchi.com](http://www.buchi.com)