

Certificate of Analysis

Kjeldahl Tablets Titanium Micro

Order No.: 11057981
Product: Kjeldahl Tablets Titanium Micro

Version created by: ODER
Version creation date: October 03, 2019
Manufacturing date: February 11, 2019
Expiry date: January 11, 2024



Lot Number 143211

Method Detail	Specification	Result
Appearance	Round white with green or blue speckled flat tablets with chalky surface	complies
Average weight	1.48 – 1.70g	1.59 g
Weight uniformity	not more than 10% outside 1.51 – 1.67 g None outside 1.47 – 1.90 g	1.57 – 1.60 g
Diameter	10 mm – 12 mm	11 mm
Friability	not more than 5%	4%
Nitrogen	not more than 0.05%	Tested every 10 th batch
Copper sulfate • 5H ₂ O	41 – 50 mg/tablet	Tested every 10 th batch

This document has been generated electronically and is valid without a signature.

BÜCHI Labortechnik AG

Meierseggstrasse 40
9230 Flawil
Switzerland

T +41 71 394 63 63
F +41 71 394 64 64
www.buchi.com