

CERTIFICATE OF ANALYSIS

PRODUCT:	Boric acid 2% with Sher indicator, pH 4.65 @ 20°C
ORDER No.:	11064972
LOT No.:	BOAIS2517K1
DATE OF TEST:	25/10/2017
PRODUCTION DATE:	10/10/2017
EXPIRY DATE:	28/04/2019
SPECIFICATION:	1.96 – 2.04% Boric Acid
MEASURED VALUE:	2.00%
TEST METHOD:	Tested by potentiometric titration, to in-house method.
TRACEABILITY:	This solution was checked by means of Sodium Hydroxide Analytical Volumetric Standard. This volumetric standard is directly traceable to a Standard Reference Material of National Institute of Standards and Technology (USA), 84L Potassium Hydrogen Phthalate. All raw materials used to prepare this product are of high purity.
SENSITIVITY TEST:	pH >7.5: Green pH 7.4 – 4.8: Blue pH 4.65: Grey / Brown

This certificate is electronically generated and does not require a signature.

Date: 25/10/2017

This certificate must not be reproduced except in full.