

CERTIFICATE OF ANALYSIS

PRODUCT: Boric Acid 2% with Sher Indicator, pH 4.65 @ 20°C

ORDER No: 11064972

LOT No: BOAIS2520A1

DATE OF TEST: 20/01/2020

PRODUCTION DATE: 17/01/2020

EXPIRY DATE: 28/07/2021

SPECIFICATION: 1.96 – 2.04% Boric Acid

- **MEASURED VALUE:** 2.03% Boric Acid
- **TEST METHOD:** Tested by potentiometric titration, to in-house method.

TRACEBILITY: This solution was checked by means of Sodium Hydroxide Analytical Volumetric Standard. This volumetric standard is directly traceable to a Standard Reference Material of National Institute of Standards and Technology (USA), 84L Potassium Hydrogen Phthalate. All raw materials used to prepare this product are of high purity.

SPECIFICATION: 4.50 – 4.80 pH @ 20°C

MEASURED VALUE: 4.63 pH @ 20°C

SENSITIVITY TEST: pH > 7.5: Green pH 7.4 - 4.8: Blue pH 4.65: Brown

This certificate is electronically generated and does not require a signature.

Date: 20/01/2020

This certificate must not be reproduced except in full.

BÜCHI Labortechnik AG Meierseggstrasse 40 9230 Flawil Switzerland T +41 71 394 63 63 F +41 71 394 64 64 www.buchi.com