

CERTIFICATE OF ANALYSIS

PRODUCT: Boric acid 2% with Sher indicator, pH 4.65 @ 20°C

ORDER No: 11064972

LOT No: BOAIS2516K1

DATE OF TEST: 25/10/2016

PRODUCTION DATE: 21/10/2016

EXPIRY DATE: 28/04/2018

SPECIFICATION: 1.96 – 2.04% Boric Acid

MEASURED VALUE: 2.00%

TEST METHOD: Tested by potentiometric titration, to in-house method.

TRACEBILITY: This solution was checked by means of Sodium Hydroxide

Analytical Volumetric Standard. This volumetric standard is directly traceable to a Standard Reference Material of National Institute of Standards and Technology (USA), 84L Potassium Hydrogen Phthalate. All raw materials used to prepare this

product are of high purity.

SENSITIVITY TEST: pH > 7.5: Green

pH 7.4 – 4.8: Blue pH 4.65: Brown

This certificate is electronically generated and does not require a signature.

Date: 25/10/2016

This certificate must not be reproduced except in full.

BÜCHI Labortechnik AG Meierseggstrasse 40 9230 Flawil Switzerland T +41 71 394 63 63 F +41 71 394 64 64 www.buchi.com