

CERTIFICATE OF ANALYSIS

PRODUCT:	Boric Acid 4% with Sher Indicator, pH 4.65 @ 20°C
ORDER No:	11064973
LOT No:	BOAIS4519H1
DATE OF TEST:	09/08/2019
PRODUCTION DATE:	08/08/2019
EXPIRY DATE:	28/02/2021
SPECIFICATION:	3.92 – 4.08% Boric Acid
MEASURED VALUE:	4.06% Boric Acid
TEST METHOD:	Tested by potentiometric titration, to in-house method.
TRACEABILITY:	This solution was checked by means of Sodium Hydroxide Analytical Volumetric Standard. This volumetric standard is directly traceable to a Standard Reference Material of National Institute of Standards and Technology (USA), 84L Potassium Hydrogen Phthalate. All raw materials used to prepare this product are of high purity.
SPECIFICATION:	4.50 – 4.80 pH @ 20°C
MEASURED VALUE:	4.60 pH @ 20°C
SENSITIVITY TEST:	pH > 7.5: Green pH 7.4 – 4.8: Blue pH 4.65: Brown

This certificate is electronically generated and does not require a signature.

Date: 02/09/2019

This certificate must not be reproduced except in full.