



BUCHI Certificate of Analysis (CoA)

Product Characteristics

Product:	pH Buffer Solution pH4.00 (Red) \pm 0.01 @ 20°C
Order code:	11064974
LOT no.:	4C23H1
Date of test:	30/08/2023
Production date:	30/08/2023
Expiration date:	28/08/2026

Testing Results

Description	Specification	Measured value
pH value (@20 °C)	3.99 – 4.01	4.00
Buffer substance	Potassium Hydrogen Phthalate	

pH Measurement:

The result reported above was determined by analysis of a sample of this lot taken at time of manufacture. Measured with a combination glass electrode after multiple point calibration with reference materials. It is certified traceable to the following National Institute of Standards and Technology (USA), SRM 185i Potassium Hydrogen Phthalate, SRM 186-I-g Potassium Dihydrogen Phosphate and SRM 186-II-g Disodium Hydrogen Phosphate.

This certificate relates solely to the lot number given above. The uncertainty of measurement has been calculated not to exceed \pm 0.01pH at 95% confidence level, i.e., coverage factor $k=2$.

Certificate issue date: 18.09.2023

This certificate is electronically generated and does not require a signature.

This certificate must not be reproduced except in full.