



BÜCHI Labortechnik AG

Meierseggrasse 40
CH-9230 Flawil
Switzerland

T +41 71 394 63 63
www.buchi.com

BUCHI Certificate of Analysis (CoA)

Product Characteristics

Product: Kjeldahl Tablets Titanium
Order code: 11072627
Order no.: 110203570
LOT no.: 04010724
Production date: 01.07.2024
Expiration date: 01.07.2029

Testing Results

Parameter	Specification	Measured value
Appearance	Round white with green/blue speckled flat tablets	Complies
Potassium sulfate	3.50 g/tablet	3.50 g/tablet
Copper sulfate pentahydrate	0.105 g/tablet	0.105 g/tablet
Titanium dioxide	0.105 g/tablet	0.105 g/tablet
Weight uniformity	3.71 ± 0.2 g	3.70 ± 0.01 g
Friability	≥ 100 N/mm ²	214 ± 19 N/mm ²
Nitrogen	≤ 12 ppm	≤ 7 ppm

Storage: Room temperature, keep dry and tightly closed in well-ventilated rooms.

Certificate issue date: 29.08.2024

This certificate is electronically generated and does not require a signature.

This certificate must not be reproduced except in full.