



BÜCHI Labortechnik AG

Meierseggrasse 40
CH-9230 Flawil
Switzerland

T +41 71 394 63 63
www.buchi.com

BUCHI Certificate of Analysis (CoA)

Product Characteristics

Product: Kjeldahl Tablets Antifoam
Order code: 11072631
Order no.: 2021-2117
LOT no.: 11020621
Production date: 02.06.2021
Expiration date: 02.06.2026

Testing Results

| Parameter | Specification | Measured value |
|-------------------|------------------------|------------------------------|
| Appearance | Round white tablets | Complies |
| Sodium sulfate | 0.97 g/tablet | 0.97 g/tablet |
| Antifoam | 0.03 g/tablet | 0.03 g/tablet |
| Weight uniformity | 1.0 ± 0.1 g | 1.01 ± 0.02 g |
| Friability | ≥ 50 N/mm ² | 60.4 ± 7.2 N/mm ² |
| Nitrogen | ≤ 12 ppm | ≤ 6 ppm |

Storage: Room temperature, keep dry and tightly closed in well-ventilated rooms.

Certificate issue date: 23.06.2021

This certificate is electronically generated and does not require a signature.

This certificate must not be reproduced except in full.