



BÜCHI Labortechnik AG

Meierseggrasse 40  
CH-9230 Flawil  
Switzerland

T +41 71 394 63 63  
www.buchi.com

## BUCHI Certificate of Analysis (CoA)

### Product Characteristics

Product:	Kjeldahl Tablets Antifoam
Order code:	11072631
Order no.:	110204447
LOT no.:	11200624
Production date:	20.06.2024
Expiration date:	20.06.2029

### Testing Results

Parameter	Specification	Measured value
Appearance	Round white tablets	Complies
Sodium sulfate	0.97 g/tablet	0.97 g/tablet
Antifoam	0.03 g/tablet	0.03 g/tablet
Weight uniformity	1.0 ± 0.1 g	1.01 ± 0.02 g
Friability	≥ 50 N/mm <sup>2</sup>	65 ± 4 N/mm <sup>2</sup>
Nitrogen	≤ 12 ppm	≤ 7 ppm

Storage: Room temperature, keep dry and tightly closed in well-ventilated rooms.

Certificate issue date: 22.01.2025

This certificate is electronically generated and does not require a signature.

This certificate must not be reproduced except in full.