



BÜCHI Labortechnik AG

Meierseggrasse 40  
CH-9230 Flawil  
Switzerland

T +41 71 394 63 63  
[www.buchi.com](http://www.buchi.com)

## BUCHI Certificate of Analysis (CoA)

### Product Characteristics

Product:	SHER Mixed Indicator
Order code:	003512
LOT no.:	SHER25J1
Date of test:	16/09/2025
Production date:	16/09/2025
Expiration date:	28/03/2027

### Testing Results

Description	Specification	Measured value
Sensitivity test with 4% Boric Acid	pH 4: Orange pH 7.5: Green	pH 4: Orange pH 7.5: Green

Certificate issue date: 17/09/2025

This certificate is electronically generated and does not require a signature.

This certificate must not be reproduced except in full.