



BÜCHI Labortechnik AG

Meierseggstrasse 40
CH-9230 Flawil
Switzerland

T +41 71 394 63 63
www.buchi.com

BUCHI Certificate of Analysis (CoA)

Product Characteristics

Product:	Kjeldahl Tablets Missouri
Order code:	11072629
Order no.:	110206379
LOT no.:	08150126
Production date:	15.01.2026
Expiration date:	15.01.2031

Testing Results

Parameter	Specification	Measured value
Appearance	Round white-blue tablets	Complies
Potassium sulfate	4.98 g/tablet	4.98 g/tablet
Copper sulfate pentahydrate	0.02 g/tablet	0.02 g/tablet
Weight uniformity	5.0 ± 0.2 g	5.0 ± 0.04 g
Friability	≥ 100 N/mm ²	428,72 ± 85 N/mm ²
Nitrogen	≤ 12 ppm	≤ 6 ppm

Storage: Room temperature, keep dry and tightly closed in well-ventilated rooms.

Certificate issue date: 27.01.2026

This certificate is electronically generated and does not require a signature.

This certificate must not be reproduced except in full.