



BÜCHI Labortechnik AG

Meierseggrasse 40
CH-9230 Flawil
Switzerland

T +41 71 394 63 63
www.buchi.com

BUCHI Certificate of Analysis (CoA)

Product Characteristics

Product:	Kjeldahl Tablets Titanium
Order code:	11072627
Order no.:	110103763
LOT no.:	04020626
Production date:	02.06.2026
Expiration date:	02.06.2031

Testing Results

Parameter	Specification	Measured value
Appearance	Round white with green/blue speckled flat tablets	Complies
Potassium sulfate	3.50 g/tablet	3.50 g/tablet
Copper sulfate pentahydrate	0.105 g/tablet	0.105 g/tablet
Titanium dioxide	0.105 g/tablet	0.105 g/tablet
Weight uniformity	3.71 ± 0.2 g	3.71 ± 0.03 g
Friability	≥ 100 N/mm ²	$231,66 \pm 53$ N/mm ²
Nitrogen	≤ 12 ppm	≤ 6 ppm

Storage: Room temperature, keep dry and tightly closed in well-ventilated rooms.

Certificate issue date: 24.06.2026

This certificate is electronically generated and does not require a signature.

This certificate must not be reproduced except in full.