



BÜCHI Labortechnik AG

Meierseggsstrasse 40
CH-9230 Flawil
Switzerland

T +41 71 394 63 63
www.buchi.com

BUCHI Certificate of Analysis (CoA)

Product Characteristics

Product:	Boric acid 2%
Order code:	11086758
LOT no.:	20261269
Production date:	27/06/2026
Expiration date:	27/12/2027

Testing Results

Description	Specification	Measured value
Boric acid content	1.96 – 2.04 %	1.99 %
pH value (@20 °C)	4.50 – 4.80	4.68

Test method: Tested by potentiometric titration, to in-house method.

Traceability: This solution was checked by means of Sodium Hydroxide Analytical Volumetric Standard. This volumetric standard is directly traceable to a Standard Reference Material of National Institute of Standards and Technology (USA), 84L Potassium Hydrogen Phthalate. All raw materials used to prepare this product are of high purity.

Storage: Room temperature, keep dry and tightly closed.

Certificate issue date: 29/07/2026

This certificate is electronically generated and does not require a signature.

This certificate must not be reproduced except in full.