



BÜCHI Labortechnik AG

Meierseggstrasse 40
CH-9230 Flawil
Switzerland

T +41 71 394 63 63
www.buchi.com

BUCHI Certificate of Analysis (CoA)

Product Characteristics

Product:	Boric acid 4%
Order code:	11086757
LOT no.:	20261550
Production date:	03/08/2026
Expiration date:	02/02/2028

Testing Results

Description	Specification	Measured value
Boric acid content	3.92 – 4.08 %	3.98 %
pH value (@20 °C)	4.50 – 4.80	4.65

Test method: Tested to in-house method.

Traceability: This solution was checked by means of Sodium Hydroxide Analytical Volumetric Standard. This volumetric standard is directly traceable to a Standard Reference Material of National Institute of Standards and Technology (USA), 84L Potassium Hydrogen Phthalate. All raw materials used to prepare this product are of high purity.

Storage: Room temperature, keep dry and tightly closed.

Certificate issue date: 17/08/2026

This certificate is electronically generated and does not require a signature.

This certificate must not be reproduced except in full.