



BÜCHI Labortechnik AG

Meierseggstrasse 40
CH-9230 Flawil
Switzerland

T +41 71 394 63 63
www.buchi.com

BUCHI Certificate of Analysis (CoA)

Product Characteristics

Product:	Kjeldahl Tablets Antifoam
Order code:	11072631
Order no.:	110103908
LOT no.:	11110526
Production date:	11.05.2026
Expiration date:	11.05.2031

Testing Results

Parameter	Specification	Measured value
Appearance	Round white tablets	Complies
Sodium sulfate	0.97 g/tablet	0.97 g/tablet
Antifoam	0.03 g/tablet	0.03 g/tablet
Weight uniformity	1.0 ± 0.1 g	1.0 ± 0.02 g
Friability	≥ 50 N/mm ²	73,8 ± 12 N/mm ²
Nitrogen	≤ 12 ppm	≤ 7 ppm

Storage: Room temperature, keep dry and tightly closed in well-ventilated rooms.

Certificate issue date: 11.08.2026

This certificate is electronically generated and does not require a signature.

This certificate must not be reproduced except in full.