



BÜCHI Labortechnik AG

Meierseggrasse 40
CH-9230 Flawil
Switzerland

T +41 71 394 63 63
www.buchi.com

BUCHI Certificate of Analysis (CoA)

Product Characteristics

Product:	pH Buffer Solution pH4.00 (Red) ± 0.01 @ 20°C
Order code:	1040060C
LOT no.:	4C26G1
Date of test:	08/07/2026
Production date:	08/07/2026
Expiration date:	28/07/2029

Testing Results

Description	Specification	Measured value
pH value (@20 °C)	3.99 – 4.01	4.00
Buffer substance	Potassium Hydrogen Phthalate	

pH Measurement:

The result reported above was determined by analysis of a sample of this lot taken at time of manufacture. Measured with a combination glass electrode after multiple point calibration with reference materials. It is certified traceable to the following National Institute of Standards and Technology (USA), SRM 185i Potassium Hydrogen Phthalate, SRM 186-I-g Potassium Dihydrogen Phosphate and SRM 186-II-g Disodium Hydrogen Phosphate.

This certificate relates solely to the lot number given above. The uncertainty of measurement has been calculated not to exceed ± 0.01 pH at 95% confidence level, i.e., coverage factor $k=2$.

Certificate issue date: 27.08.2026

This certificate is electronically generated and does not require a signature.

This certificate must not be reproduced except in full.