



BÜCHI Labortechnik AG

Meierseggrasse 40
CH-9230 Flawil
Switzerland

T +41 71 394 63 63
www.buchi.com

BUCHI Certificate of Analysis (CoA)

Product Characteristics

Product:	Saturated Potassium Chloride (KCl), free from Silver Ion
Product no.:	LKCL060
LOT no.:	LKCL26E1
Testing date:	15.05.2026
Production date:	15.05.2026
Expiry date:	28.05.2029

Testing Results

Description	Specification	Measured value
Potassium Chloride content at @ 20 °C	3.96 - 4.04 M KCl	3.97 M KCl

Method:

The result reported above was determined by analysis of a sample of this lot taken at time of manufacture. This Solution was tested by precipitation titration with Silver Nitrate Analytical Volumetric Standard.

Metrological Traceability:

The test result is directly traceable to Standard Reference Material of National Institute of Standards and Technology (USA), SRM 919b Sodium Chloride.

Certificate issue date: 27.08.2026

This certificate is electronically generated and does not require a signature.

This certificate must not be reproduced except in full.